
Nmr Chemical Shifts Of Common Laboratory Solvents As Trace Impurities

*Nmr Chemical Shifts Of
Common Laboratory
Solvents As Trace
Impurities*

Downloaded from
web1.enteraskincare.com
by May & Oconnell

INTRODUCTION: THE HIDDEN SIGNATURES IN YOUR NMR SPECTRUM

Nuclear Magnetic Resonance (NMR) spectroscopy stands as one of the most powerful and versatile analytical tools in the chemist's arsenal. Whether you are elucidating the structure of a novel natural product, confirming the identity of a synthetic intermediate, or assessing the purity of a final compound, the NMR spectrum offers a wealth of information. However, alongside the peaks of your target molecule, a spectrum almost invariably contains signals from substances that were not intentionally added. Among the most common and persistent of these unwelcome guests are trace amounts of the laboratory solvents used during synthesis, workup, and purification.

Understanding the **nmr chemical shifts of common laboratory solvents as trace impurities** is not merely an exercise in academic curiosity; it is a fundamental skill for every practicing chemist. Misidentifying a residual solvent peak can lead to erroneous structural assignments, inflated purity claims, and wasted hours of troubleshooting. This article provides a

comprehensive guide to recognizing, interpreting, and managing these ubiquitous spectral artifacts, empowering you to read your NMR spectra with greater confidence and precision.

WHAT ARE NMR CHEMICAL SHIFTS AND WHY DO SOLVENT IMPURITIES APPEAR?

Before diving into specific values, it is essential to grasp the physical phenomenon at play. A chemical shift, reported in parts per million (ppm), describes the resonance frequency of a nucleus relative to a standard reference, typically tetramethylsilane (TMS) for ^1H and ^{13}C NMR. This frequency is exquisitely sensitive to the electronic environment surrounding the nucleus.

Solvent impurities appear in NMR spectra for a simple reason: no purification method is perfect. A small amount of the solvent used for a reaction or extraction can remain trapped in the final product, especially if it has a high boiling point or forms a stable solvate. During workup, solvents like ethyl acetate, dichloromethane, or diethyl ether are frequently used and can persist as trace contaminants. When this slightly impure sample is dissolved in a deuterated NMR solvent (such as CDCl_3 , $\text{DMSO}-d_6$, or CD_3OD), the residual non-deuterated solvent molecules

become visible as sharp, characteristic peaks in the spectrum.

Recognizing these peaks is critical. The **nmr chemical shifts of common laboratory solvents as trace impurities** follow predictable patterns, but they can vary slightly depending on the concentration, the deuterated solvent used, and the temperature of the experiment.

THE MOST FREQUENT CULPRITS: A DEEP DIVE INTO COMMON SOLVENT PEAKS

Proton NMR (^1H NMR) in Deuterated Chloroform (CDCl_3)

Deuterated chloroform is the most widely used NMR solvent, and a substantial body of reference data exists for impurity peaks in this medium. When examining a ^1H NMR spectrum acquired in CDCl_3 , several solvent impurities are routinely encountered.

- **Water (H_2O):** Perhaps the most universal impurity. In CDCl_3 , residual water typically appears as a broad singlet around **1.5 - 1.6 ppm**. The exact shift is highly concentration-dependent; in very dry samples, it can move upfield, while in wet samples, it broadens and shifts downfield. Do not confuse this with a hydroxyl or amine proton from your

compound.

- **Dichloromethane (CH_2Cl_2):** A workhorse solvent for extractions and chromatography. It gives a sharp, unmistakable singlet at approximately **5.30 ppm** in CDCl_3 .
- **Ethyl Acetate (EtOAc):** Extremely common after column chromatography. Its characteristic pattern includes a quartet for the OCH_2CH_3 group near **4.12 ppm**, a sharp singlet for the $\text{CH}_3\text{C}=\text{O}$ group at **2.05 ppm**, and a triplet for the OCH_2CH_3 group around **1.26 ppm**. The methyl singlet at 2.05 ppm is often the most diagnostic and least overlapped signal.
- **Diethyl Ether (Et_2O):** Another frequent contaminant from extractions. It displays a quartet for the OCH_2CH_3 groups near **3.41 ppm** and a triplet for the OCH_2CH_3 groups around **1.21 ppm**.
- **Acetone ($(\text{CH}_3)_2\text{CO}$):** Used for rinsing glassware and in some purification protocols. Its residual peak is a singlet at approximately **2.17 ppm** in CDCl_3 .
- **Acetonitrile (CH_3CN):** A common HPLC solvent. It gives a sharp singlet for the methyl group at roughly **2.10 ppm** in CDCl_3 , very close to the acetone peak and the ethyl acetate methyl peak. Careful inspection of coupling patterns is needed to differentiate them.
- **Dimethyl Sulfoxide (DMSO):** Often used as a reaction solvent but difficult to remove completely. In CDCl_3 , residual DMSO appears as a singlet at approximately **2.62 ppm**.
- **Methanol (CH_3OH):** A common protic solvent. In CDCl_3 , the methyl

group appears as a singlet at about **3.49 ppm**, and the hydroxyl proton appears as a broad, often featureless peak that can vary from **1.5 to 2.0 ppm** depending on concentration and temperature.

- **n-Hexane / Petroleum Ether:** These are complex mixtures of alkanes. They appear as a series of overlapping, often unresolved multiplets in the aliphatic region (roughly **0.8 - 1.6 ppm**), making them difficult to distinguish from your compound's alkyl chain signals.

Proton NMR in Other Deuterated Solvents

The **nmr chemical shifts of common laboratory solvents as trace impurities** are not fixed; they shift depending on the deuterated solvent in which the sample is dissolved. Being aware of these solvent-dependent changes is crucial.

- **In DMSO-*d*₆:** The residual water peak shifts downfield significantly, often appearing between **3.3 and 3.4 ppm**, which can obscure signals from NH₂ or OH groups. Dichloromethane shifts to about **5.76 ppm**, ethyl acetate signals appear at approximately **4.03 ppm** (q), **1.99 ppm** (s), and **1.17 ppm** (t), and acetone appears at **2.09 ppm**.
- **In Deuterated Methanol**

(CD₃OD): The residual water and methanol hydroxyl peaks exchange rapidly, forming a combined broad peak often around **4.8 - 4.9 ppm**. This can mask the region where many anomeric protons appear. Acetone in CD₃OD is found near **2.15 ppm**, and dichloromethane near **5.49 ppm**.

- **In Deuterated Acetone (Acetone-*d*₆):** Water appears around **2.8 - 2.9 ppm**. Dichloromethane is at approximately **5.63 ppm**, and ethyl acetate signals are near **4.05 ppm** (q), **1.97 ppm** (s), and **1.20 ppm** (t).

CARBON NMR (¹³C NMR) IMPURITY PEAKS

While ¹H NMR is more sensitive and often used for purity assessment, ¹³C NMR spectra also contain solvent impurity peaks. These are generally less intense due to the lower natural abundance of ¹³C, but they can be informative, especially when ¹H signals are crowded.

- **Dichloromethane:** A single peak near **53.1 - 54.0 ppm** in most solvents.
- **Ethyl Acetate:** Peaks at approximately **170.5 ppm** (C=O), **60.2 ppm** (OCH₂), **20.5 ppm** (CH₃C=O), and **14.1 ppm** (CH₃CH₂).
- **Acetone:** Peaks at about **205.0 ppm** (C=O) and **30.5 ppm** (CH₃).
- **Acetonitrile:** Peaks at roughly **117.7 ppm** (CN) and **1.7 ppm** (CH₃).
- **Dimethyl Sulfoxide:** Peaks at about **40.5 ppm** (CH₃).
- **Diethyl Ether:** Peaks at about

66.0 ppm (OCH_2) and **15.0 ppm** (CH_3).

- **Methanol:** Peaks at about **49.5 ppm** (CH_3).

FACTORS THAT INFLUENCE THE CHEMICAL SHIFTS OF IMPURITIES

Relying solely on memorized values can lead to misidentification. Several factors can cause the **nmr chemical shifts of common laboratory solvents as trace impurities** to deviate from textbook values.

Concentration

Dependence

This is most pronounced for exchangeable protons, such as those in water, methanol, and other alcohols. As the concentration of the impurity increases, intermolecular hydrogen bonding changes the electronic shielding, causing the peak to shift downfield. A trace amount of water in CDCl_3 might appear at 1.5 ppm, but a more substantial amount could shift to 1.8 ppm or higher.

Temperature

NMR experiments are often run