
One Touch Ultra 2 Control Solution

One Touch Ultra 2 Control Solution

Downloaded from
web1.enteraskincare.com by Gillian &
Dorsey

INTRODUCTION: WHY ACCURACY MATTERS IN BLOOD GLUCOSE MONITORING

For millions of people managing diabetes, the blood glucose meter is an indispensable daily tool. A simple finger prick provides a number that dictates dietary choices, insulin dosages, and physical activity. But what happens when that number is not quite right? Inaccurate readings can lead to incorrect insulin doses, dangerous hypoglycemia, or unchecked hyperglycemia. This is where the **One Touch Ultra 2 Control Solution** enters the picture. It is not merely an accessory; it is a critical component of a reliable diabetes management routine. This article explores everything you need to know about this control solution, from its purpose and proper usage to interpreting results and ensuring your meter is always performing at its best.

WHAT IS ONE TOUCH ULTRA 2 CONTROL SOLUTION?

The **One Touch Ultra 2 Control Solution** is a liquid testing medium specifically formulated to verify that your One Touch Ultra 2 meter and test strips are working correctly together. It contains a known amount of glucose, which allows you to check

the accuracy of the entire testing system. Think of it as a calibration check. Just as you would test a thermometer against a known temperature, this control solution provides a benchmark for your glucose monitor.

It is important to distinguish this from the blood samples you take daily. The control solution is not for human consumption or injection. It is strictly for external, diagnostic use. The solution is designed to mimic a blood sample in its interaction with the test strip, ensuring that the meter's optics and electronics are functioning within an acceptable range. Using a **One Touch Ultra 2 Control Solution** test regularly helps you catch potential issues before they affect your real-world readings.

Who Should Use It?

While any user of the One Touch Ultra 2 meter can benefit from regular testing, the control solution is especially valuable for:

- **New users** who are learning how to use the meter and strips for the first time.
- **Those experiencing inconsistent readings** that do not match how they feel.
- **Users who have dropped or damaged the meter** and

want to ensure it still works.

- **Anyone starting a new vial of test strips** to confirm they are not defective.
- **Individuals who store strips in extreme temperatures** or high humidity.

WHY CONTROL SOLUTION TESTING IS MORE IMPORTANT THAN YOU THINK

Many people skip control solution testing because they assume their meter is always correct. However, several factors can compromise accuracy. Test strips are sensitive to environmental conditions like heat, moisture, and light. An expired strip or a strip exposed to humidity can give a false reading. The meter itself can accumulate dust or damage over time. By using the **One Touch Ultra 2 Control Solution**, you are performing a quality assurance check on your entire testing system.

Consider this scenario: you test your blood sugar before a meal and get a reading of 180 mg/dL. You take your insulin accordingly. But what if the actual value was 130 mg/dL? You would be taking too much insulin, potentially leading to a low blood sugar event later. A simple control solution test could have flagged the discrepancy. For people on intensive insulin therapy, even small inaccuracies can have significant consequences. Regular use of the control solution provides peace of mind and reinforces the reliability of your self-monitoring.

HOW TO USE THE ONE TOUCH ULTRA 2 CONTROL

SOLUTION: A STEP-BY-STEP GUIDE

Using the control solution is straightforward, but it must be done correctly to obtain a valid result. Follow these steps carefully.

Step 1: Gather Your Supplies

You will need your One Touch Ultra 2 meter, a test strip, and the **One Touch Ultra 2 Control Solution** bottle. Ensure the control solution is at room temperature before testing. Cold solution can affect the reading.

Step 2: Check the Expiration Date

Never use an expired control solution. The bottle has a printed expiration date. Also, note that once opened, the solution is typically good for 90 days. Write the discard date on the bottle label when you first open it.

Step 3: Insert a Test Strip

Insert a new test strip into your meter. The meter should turn on automatically and display a code number. Ensure the code on the meter screen matches the code on the test strip vial. If not, follow

your meter manual to change the code.

Step 4: Prepare the Control Solution

Shake the control solution bottle vigorously. This ensures the glucose is evenly distributed. Then, squeeze the bottle gently to discard the first drop. This drop can be less concentrated. The second drop is the one you want to test.

Step 5: Apply the Solution to the Strip

Touch the tip of the test strip to the drop of control solution. The strip will draw the solution in by capillary action. Hold it there until the meter beeps or the display confirms enough solution has been applied.

Step 6: Read and Compare the Result

The meter will display a number after a few seconds. This is your control solution result. Compare this number to the range printed on the test strip vial or the control solution bottle. The range is clearly labeled, for example, "Control Solution Range: 80-120

mg/dL."

Step 7: Interpret the Result

- **Result is within range:** Your meter and strips are working correctly. You can proceed with regular testing.
- **Result is out of range:** Do not use the meter for blood glucose testing until you resolve the issue. Try again with a new test strip. If it is still out of range, check the expiration dates of both the strips and the control solution. Ensure the solution was shaken well and at room temperature. If the problem persists, contact the manufacturer or consider replacing the meter.

INTERPRETING CONTROL SOLUTION RESULTS

The key to effective quality control is understanding what the result means. The **One Touch Ultra 2 Control Solution** has a printed target range. This is not a narrow single number but a range, usually spanning about 20-30 mg/dL. For example, you might see "Range: 85-115 mg/dL." Your result must fall within this entire range to pass.

It is also important to understand that control solution results are not the same as blood glucose results. Do not use the control solution number to adjust your medication or food intake. It is purely a system check. If your control solution test passes, you can trust that your blood glucose readings are likely accurate. If it fails, troubleshoot immediately. Never assume a failing control solution test is irrelevant. It is a warning sign that something is

wrong with your testing system.

STORAGE AND SHELF LIFE OF THE CONTROL SOLUTION

Proper storage of the **One Touch Ultra 2 Control Solution** is essential for maintaining its integrity. Like test strips, the solution is sensitive to environmental factors.

Storage Guidelines

- **Temperature:** Store the bottle at room temperature, typically between 36°F and 86°F (2°C to 30°C). Avoid extreme heat, cold, or direct sunlight.
- **Cap Tightly Closed:** Always replace the cap immediately after use. Exposure to air can alter the glucose concentration.
- **Clean Tip:** Keep the bottle tip clean. Do not touch it with your fingers or the test strip.
- **Do Not Refrigerate:** While it might seem logical to store it cold, refrigeration can cause condensation when the bottle is opened, potentially contaminating the solution.

Shelf Life

Unopened, the solution is stable until the printed expiration date. Once opened, it is generally usable for 90 days. Mark the date of opening on the bottle so you remember when to discard it. Using old or improperly stored solution can lead to inaccurate control

tests, which may cause you to mistakenly believe your meter is faulty.

COMMON QUESTIONS ABOUT THE ONE TOUCH ULTRA 2 CONTROL SOLUTION

Users often have questions about the practical aspects of control solution testing. Here are answers to some of the most common inquiries.

How Often Should I Test?

There is no universal rule, but best practices suggest testing:

- When you open a new vial of test strips.
- If you suspect the meter or strips have been damaged.
- Periodically, such as once a week, to ensure ongoing accuracy.
- Whenever you get a blood glucose result that seems unusually high or low compared to your symptoms.

What If I Use the Wrong Control Solution?

Control solutions are specific to meter brands and models. Using a different brand or type of solution may not give accurate results. Always use the **One Touch Ultra 2 Control Solution** for your specific meter. The chemistry of the solution is designed

to work with the unique calibration of the One Touch Ultra 2 system.

Can I Use the Control Solution With Other Meters?

No. The **One Touch Ultra 2 Control Solution** is formulated for the One Touch Ultra 2 system. Using it with other meters will not provide a reliable quality check. Each meter brand has its own control solution with a specific glucose concentration and range.

WHERE TO BUY AND WHAT TO LOOK FOR

Finding the **One Touch Ultra 2 Control Solution** is generally easy. It is available at most pharmacies, including major chains, and through online retailers. When purchasing, you should consider a few factors to ensure you get a genuine, effective product.

Authenticity and Expiration

Always buy from a reputable source. Check the expiration date on the bottle before purchasing. Avoid buying from third-party sellers who may not store the product properly. An expired or improperly stored control solution is useless for quality assurance.

Bottle Size

The control solution typically comes in a small bottle containing 2.5 mL or 5 mL of solution. A single bottle contains enough for multiple tests. Given the 90-day discard window after opening, a standard bottle is more than sufficient for most users.

Price and Availability

The cost is generally modest, especially compared to the cost of test strips. Some insurance plans may cover control solution, but it is often an out-of-pocket expense. Check with your pharmacy or supplier for pricing. In many cases, it is a small investment that protects the accuracy of your much more expensive test strips.

TROUBLESHOOTING: WHEN YOUR CONTROL SOLUTION TEST FAILS

A failed control solution test can be frustrating, but it is a solvable problem. Here is a systematic approach to troubleshooting.

1. **Check the expiration date** of the control solution and the test strips. Replace any that are expired.
2. **Ensure proper storage.** Was the control solution left in a hot car or exposed to humidity?
3. **Shake the bottle thoroughly.** The solution must be well-mixed before each use.
4. **Discard the first drop.** Always use the second drop for

testing.

5. **Use a fresh test strip.** Sometimes a single strip is defective.
6. **Verify the code.** Make sure the meter code matches the test strip vial code.
7. **Clean the meter.** Dust or residue on the meter's optical window can affect readings.
8. **Test again at room temperature.** Cold or hot conditions can skew results.

If you have tried all these steps and the test still fails, contact the manufacturer's customer support. The meter itself may need to

be replaced. Do not continue using a meter that consistently fails a control solution test with fresh, properly stored supplies.

THE ROLE OF CONTROL SOLUTION IN OVERALL DIABETES MANAGEMENT

Accurate blood glucose data is the foundation of effective diabetes management. The **One Touch Ultra 2 Control Solution** is a simple tool that ensures this foundation is solid. By incorporating regular control solution testing into your routine, you are taking an active step toward reliable self-monitoring. This is particularly important for individuals who rely on their meter data to make