
Hill Rom Medical Gas Design Guide

*Hill Rom Medical Gas
Design Guide*

*Downloaded from
web1.enteraskincare.com
by Jennings &
McConnell*

INTRODUCTION TO THE HILL ROM MEDICAL GAS DESIGN GUIDE

Medical gas systems are the silent lifelines of modern healthcare facilities. From supplying oxygen to patient rooms to powering surgical tools with medical air, these networks must function flawlessly under the most demanding conditions. The **Hill Rom Medical Gas Design Guide** has long served as an

authoritative resource for architects, engineers, facility managers, and healthcare planners who need to create safe, compliant, and efficient gas delivery systems. While Hillrom (now part of Baxter's product portfolio) is best known for its patient support surfaces and hospital beds, its expertise in medical gas infrastructure is equally vital. This article explores the essential components of the design guide, its role in promoting patient safety, and the practical steps it outlines for designing a robust medical gas system.

WHAT IS A MEDICAL GAS SYSTEM?

Gas Design Guide means recognizing that a well-planned gas system is not just about pipes and outlets. It is about integrating mechanical engineering with stringent healthcare regulations, ensuring reliable performance during emergencies, and preparing for future expansions. This guide helps professionals avoid common pitfalls such as undersized pipelines, improper zone valve placement, or non-compliant alarm configurations. Whether you are designing a new hospital wing or upgrading an existing facility, following the guide's recommendations can reduce risk, simplify maintenance, and improve overall clinical outcomes.

AND WHY DOES DESIGN MATTER?

A medical gas system delivers essential gases – oxygen, nitrous oxide, medical air, carbon dioxide, and vacuum – from central supply sources to point-of-use outlets in patient care areas. It also includes waste anesthesia gas disposal (WAGD) and nitrogen for surgical tools. The **Hill Rom Medical Gas Design Guide** focuses on the layout, sizing, material selection, and safety features that make such systems reliable. Poor design can lead to pressure drops, contamination, or even catastrophic failures that endanger lives. Therefore, the guide emphasizes adherence to standards like NFPA 99 (Health Care Facilities Code) and ISO 7396-1.

A central theme in the design guide is redundancy. For critical gases such as oxygen, the system must have dual source supplies, automatic changeover manifolds, and backup reserves. The guide also specifies how to calculate peak demand based on the number of beds, types of procedures, and simultaneous usage patterns. For example, an intensive care unit will have higher oxygen and vacuum demands than a general ward. By following the **Hill Rom Medical Gas Design Guide**, designers can accurately size pipelines, manifolds, and alarm systems to meet real-world clinical needs.

Key Components Defined in the Guide

- **Source Equipment:** Manifolds, bulk cryogenic tanks, medical air compressors, vacuum pumps, and associated purification systems.
- **Distribution Piping:** Copper or stainless steel tubing (typically Type K or L copper per ASTM B819) with brazed or welded joints to prevent leaks.
- **Zone Valve Boxes:** Allow isolation of gas supply to specific areas (e.g., an operating room or a

nursing unit) without shutting down the entire facility.

- **Area Alarm Panels:** Provide visual and audible warnings when pressures deviate from normal ranges in a specific zone.
- **Master Alarm Panels:** Monitor the entire system from a central location, alerting staff to source failures, low reserves, or line faults.
- **Patient Outlets:** Quick-connect terminals that accept different gas connectors (color-coded and keyed to prevent misconnections).

Each component must be selected and placed according to the **Hill Rom Medical Gas Design Guide** to ensure compatibility with existing infrastructure

and ease of future upgrades.

NAVIGATING COMPLIANCE: NFPA 99 AND BEYOND

The **Hill Rom Medical Gas Design Guide** aligns closely with NFPA 99, the primary standard governing health care facilities in the United States. This standard classifies areas by risk (Category 1 for critical spaces like ORs and ICUs, Category 2 for general patient care, Category 3 for support spaces) and imposes different requirements for gas systems. The guide provides detailed tables and examples to help designers determine which category applies, and then lists corresponding design criteria such as pipe sizing, number of outlets, and alarm monitoring.

In addition to NFPA 99, the guide often

references CSA Z7396.1 (Canadian standard) and international ISO standards. For facilities that serve research or specialized clinical functions, the guide may also recommend following **Hillrom-specific best practices** gained from decades of field experience. For instance, it offers guidance on locating oxygen outlets away from heat sources and avoiding sharp bends in piping that could impede flow. The **Hill Rom Medical Gas Design Guide** also dedicates sections to labeling, signage, and documentation – critical for both commissioning and long-term maintenance.

Pressure and Flow Considerations

One of the most practical sections in the guide covers pressure requirements. Oxygen is typically delivered at 50–55 psi, medical air at 50–55 psi, and vacuum at 15–25 inHg. But these nominal values must be adjusted for pipe friction losses, elevation changes, and the number of outlets in simultaneous use. The guide includes formulas and example calculations that allow engineers to compute total pressure drop from source to farthest outlet. This ensures that even during

peak demand, all outlets receive adequate flow.

The guide also warns against using components rated only for domestic water or compressed air. Medical gas fittings must be cleaned for oxygen service (free of oil and grease) and comply with ASTM F1970. Following the **Hill Rom Medical Gas Design Guide** helps avoid costly mistakes such as installing standard brass ball valves instead of listed medical gas valves, which could introduce contamination.

DESIGNING FOR PATIENT SAFETY AND OPERATIONAL EFFICIENCY

Every decision in the design phase affects how safely and efficiently a hospital can deliver care. The **Hill Rom Medical Gas Design Guide**

emphasizes the importance of physical layout. For example, zone valve boxes should be located in corridors that are accessible during emergencies but not in areas where they can be blocked by equipment. Similarly, area alarm panels should be placed in nursing stations or other constantly staffed locations.

Outlets themselves need careful positioning. In patient rooms, the guide recommends installing outlets on the headwall at 48–60 inches above the floor, allowing easy access for respiratory therapists and nurses without interfering with bed movement. For operating rooms, outlets may be placed on ceiling booms to keep the floor clear. The guide also addresses the integration of Hillrom's own bed-mounted gas rails and modular headwalls, which combine

electrical, data, and gas connections into a single unit. By following the **Hill Rom Medical Gas Design Guide**, facilities can achieve a streamlined appearance and reduced clutter.

Future-Proofing the System

Healthcare is dynamic. The guide advises designing gas systems with capacity for at least 20–30% future growth. This means installing larger-diameter mains or extra spare conduits, even if they are initially capped. The **Hill Rom Medical Gas Design Guide** also suggests using modular assembly techniques – such as prefabricated headwalls – that can be

easily swapped or expanded as clinical needs change. Planning for future additions reduces disruption and cost when a hospital adds a new wing or converts a general ward into an ICU.

INSTALLATION AND COMMISSIONING: TURNING DESIGN INTO REALITY

Even the best design will fail if installation is sloppy. The **Hill Rom Medical Gas Design Guide** dedicates significant attention to installation procedures. It specifies brazing techniques using a minimum of 15% silver filler metal to avoid embrittlement. It also mandates that all piping be purged with nitrogen before brazing to prevent oxidation on the inside of the joints. After installation, every joint must

be visually inspected and pressure-tested.

Commissioning is a multi-step process: initial pressure test (1.5 times the normal working pressure), leak test with a non-reactive gas, and then a final test with the actual gas. The guide requires blow-down to remove debris and a purity verification to ensure that oxygen content meets medical standards ($\geq 99.0\%$ for oxygen systems). The **Hill Rom Medical Gas Design Guide** recommends using a third-party testing agency to certify the system before it is put into service. Documentation of all tests, along with as-built drawings and valve numbering, must be kept for the life of the system.

Common Installation Pitfalls to Avoid

1. Using flux for brazing – flux is prohibited because it can contaminate the gas stream and cannot be removed from the interior of the pipe.
2. Installing shut-off valves that are not specifically listed for medical gas service (look for the “medical gas” certification mark).
3. Running gas piping in the same chase as electrical conduits without proper separation

(minimum 6 inches or shielding is required).

4. Failing to provide adequate access panels for future maintenance of zone valves and alarms.

The guide spells out these details so that contractors and inspectors have a clear reference. Following the **Hill Rom Medical Gas Design Guide** during installation drastically reduces the likelihood of rework or system failure.

ONGOING MAINTENANCE AND COMPLIANCE

A medical gas system is not a “set-and-forget” infrastructure. The **Hill Rom Medical Gas Design Guide** outlines a maintenance schedule that includes daily, monthly, and annual

checks. Daily tasks include verifying that audible alarms are functional and that source pressures are within normal ranges. Monthly checks involve inspecting area alarm panels and testing zone valve boxes by cycling them open and closed. Annually, a full inspection of all pipelines, joints, and outlets should be performed, along with a purity analysis of each gas type.

Documentation is crucial for accreditation by The Joint Commission or other regulatory bodies. The guide provides templates for log sheets that record all maintenance activities, test results, and any repairs. By adhering to the **Hill Rom Medical Gas Design Guide**, facility managers can quickly demonstrate compliance during surveys and audits. Moreover, a well-maintained

system reduces downtime and prevents dangerous conditions such as cross-connection or bacterial contamination in vacuum lines.

The Role of Training

No matter how well the system is designed, it requires trained operators. The guide recommends that a hospital appoint a medical gas system manager who is responsible for overseeing daily operations, coordinating with external service providers, and ensuring that clinical staff understand how to use outlets properly. The **Hill Rom Medical Gas Design Guide** includes a section on training materials and suggested

drills, such as simulating a source failure and verifying proper changeover.

CONCLUSION: A FOUNDATION FOR SAFE PATIENT CARE

The **Hill Rom Medical Gas Design Guide** is more than a technical manual – it is a blueprint for patient safety and operational excellence. By following its recommendations, healthcare facilities can design, install, and maintain medical gas systems that meet rigorous standards while adapting to future demands. From the initial calculation of peak flow rates to the final commissioning test, every step in the guide is intended to eliminate risk and ensure that life-saving gases reach patients reliably. Whether you are constructing a new hospital or

renovating an existing one, consulting the Hill Rom Medical Gas Design Guide is an investment in quality, compliance, and peace of mind. Ultimately, a properly designed medical gas system saves lives – and that is the most important outcome of all.